

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
 Data File : VX022795.D
 Acq On : 21 Jun 2021 15:25
 Operator : JC/MD
 Sample : M2783-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D2-20210617

Quant Time: Jun 22 01:13:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	54815	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	111730	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	102944	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36712	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52741	53.499	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.000%
35) Dibromofluoromethane	5.398	113	36090	50.627	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.260%
50) Toluene-d8	8.653	98	150198	53.448	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.900%
62) 4-Bromofluorobenzene	11.086	95	51112	48.092	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.180%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	13685	20.396	ug/l	88
16) Acetone	2.386	43	1736	3.933	ug/l	92
27) cis-1,2-Dichloroethene	4.507	96	1767	1.967	ug/l	95
44) Trichloroethene	7.135	130	88859	114.288	ug/l	95
64) Tetrachloroethene	9.281	164	9600	14.746	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
Data File : VX022795.D
Acq On : 21 Jun 2021 15:25
Operator : JC/MD
Sample : M2783-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D2-20210617

Quant Time: Jun 22 01:13:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

