

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
 Data File : VX022796.D
 Acq On : 21 Jun 2021 15:48
 Operator : JC/MD
 Sample : M2783-14 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP03-20210617

Quant Time: Jun 22 01:13:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	52903	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	105733	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	93100	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	30219	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50373	52.944	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.880%
35) Dibromofluoromethane	5.397	113	35626	52.810	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.620%
50) Toluene-d8	8.653	98	144236	54.238	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.480%
62) 4-Bromofluorobenzene	11.085	95	44294	44.040	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.080%
Target Compounds						
						Qvalue
44) Trichloroethene	7.129	130	7919	10.763	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
Data File : VX022796.D
Acq On : 21 Jun 2021 15:48
Operator : JC/MD
Sample : M2783-14 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP03-20210617

Quant Time: Jun 22 01:13:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

