

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
 Data File : VX022805.D
 Acq On : 21 Jun 2021 19:21
 Operator : JC/MD
 Sample : M2783-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20210617

Quant Time: Jun 22 01:16:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53696	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	108301	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97804	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36336	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	50757	52.559	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.120%	
35) Dibromofluoromethane	5.397	113	36321	52.564	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.120%	
50) Toluene-d8	8.653	98	144765	53.146	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.300%	
62) 4-Bromofluorobenzene	11.085	95	48906	47.473	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.940%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	385684	586.796	ug/l	90
12) 1,1-Dichloroethene	2.325	96	2860	4.387	ug/l	96
16) Acetone	2.386	43	1277	2.953	ug/l	91
27) cis-1,2-Dichloroethene	4.495	96	1575	1.789	ug/l	97
44) Trichloroethene	7.135	130	126373	167.684	ug/l	94
64) Tetrachloroethene	9.275	164	31388	50.748	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
Data File : VX022805.D
Acq On : 21 Jun 2021 19:21
Operator : JC/MD
Sample : M2783-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20210617

Quant Time: Jun 22 01:16:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

