

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
 Data File : VX022806.D
 Acq On : 21 Jun 2021 19:44
 Operator : JC/MD
 Sample : M2783-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D4-20210617

Quant Time: Jun 22 01:16:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	57264	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	117109	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	104803	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38864	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	54796	53.206	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.420%	
35) Dibromofluoromethane	5.403	113	38747	51.857	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.720%	
50) Toluene-d8	8.653	98	152137	51.652	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.300%	
62) 4-Bromofluorobenzene	11.085	95	52264	46.917	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.840%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	205271	292.849	ug/l	90
12) 1,1-Dichloroethene	2.325	96	1500	2.158	ug/l	90
16) Acetone	2.386	43	1250	2.711	ug/l	# 84
27) cis-1,2-Dichloroethene	4.501	96	1158	1.234	ug/l	86
44) Trichloroethene	7.129	130	26754	32.830	ug/l	96
64) Tetrachloroethene	9.275	164	10741	16.206	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062121\
Data File : VX022806.D
Acq On : 21 Jun 2021 19:44
Operator : JC/MD
Sample : M2783-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D4-20210617

Quant Time: Jun 22 01:16:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

