

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062122\
 Data File : VX029658.D
 Acq On : 21 Jun 2022 13:14
 Operator : JC/MD
 Sample : VX0621WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0621WBL01

Quant Time: Jun 22 01:10:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	265972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	441310	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	383550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	166683	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	146520	41.616	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.240%
35) Dibromofluoromethane	5.391	113	130943	45.530	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.060%
50) Toluene-d8	8.647	98	497974	47.397	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.800%
62) 4-Bromofluorobenzene	11.079	95	176399	44.522	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062122\
Data File : VX029658.D
Acq On : 21 Jun 2022 13:14
Operator : JC/MD
Sample : VX0621WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0621WBL01

Quant Time: Jun 22 01:10:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

