

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062122\
 Data File : VX029666.D
 Acq On : 21 Jun 2022 16:33
 Operator : JC/MD
 Sample : N3339-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Jun 22 01:11:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	248600	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	426016	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	366093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	157754	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	141903	43.121	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.240%
35) Dibromofluoromethane	5.391	113	125586	45.235	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.480%
50) Toluene-d8	8.653	98	478256	47.155	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.300%
62) 4-Bromofluorobenzene	11.085	95	168441	44.039	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.080%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	2401	1.830	ug/l	95
20) Methylene Chloride	2.788	84	1065	0.365	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062122\
Data File : VX029666.D
Acq On : 21 Jun 2022 16:33
Operator : JC/MD
Sample : N3339-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Jun 22 01:11:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

