

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062122\
 Data File : VX029668.D
 Acq On : 21 Jun 2022 17:42
 Operator : JC/MD
 Sample : N3290-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-33

Quant Time: Jun 22 01:12:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	246556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	416936	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	356395	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	151890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140249	42.972	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.940%
35) Dibromofluoromethane	5.391	113	123185	45.337	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.680%
50) Toluene-d8	8.653	98	471587	47.510	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.020%
62) 4-Bromofluorobenzene	11.085	95	167219	44.672	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.340%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.965	59	89207	120.637	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	684937	78.984	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062122\
Data File : VX029668.D
Acq On : 21 Jun 2022 17:42
Operator : JC/MD
Sample : N3290-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-33

Quant Time: Jun 22 01:12:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

