

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062122\
 Data File : VX029669.D
 Acq On : 21 Jun 2022 18:06
 Operator : JC/MD
 Sample : N3342-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-TB-20220610

Quant Time: Jun 22 01:12:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	245727	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	416448	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	358927	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	151084	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	141084	43.374	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.740%
35) Dibromofluoromethane	5.391	113	125205	46.134	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.260%
50) Toluene-d8	8.653	98	471578	47.564	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.120%
62) 4-Bromofluorobenzene	11.085	95	163251	43.663	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.320%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2734	2.109	ug/l	96
20) Methylene Chloride	2.794	84	864	0.299	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062122\
Data File : VX029669.D
Acq On : 21 Jun 2022 18:06
Operator : JC/MD
Sample : N3342-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-TB-20220610

Quant Time: Jun 22 01:12:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

