

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062122\
 Data File : VX029676.D
 Acq On : 21 Jun 2022 20:49
 Operator : JC/MD
 Sample : N3325-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Jun 22 01:13:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	252282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	428687	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	371014	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158349	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140224	41.989	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.980%
35) Dibromofluoromethane	5.391	113	125447	44.904	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.800%
50) Toluene-d8	8.653	98	484787	47.501	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.000%
62) 4-Bromofluorobenzene	11.086	95	169757	44.107	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.220%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	6733	5.058	ug/l	97
17) Carbon Disulfide	2.514	76	7824	1.181	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	3765	0.378	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062122\
Data File : VX029676.D
Acq On : 21 Jun 2022 20:49
Operator : JC/MD
Sample : N3325-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Jun 22 01:13:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

