

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062123\
 Data File : VX036263.D
 Acq On : 21 Jun 2023 12:14
 Operator : JC/MD
 Sample : 03296-12
 Misc : 15.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOTAL-VOC-R-01-GRAB-7

Quant Time: Jun 22 05:00:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	184455	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	335711	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	288525	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	124661	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	130356	38.672	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	77.340%#
35) Dibromofluoromethane	5.385	113	82384	36.189	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	72.380%#
50) Toluene-d8	8.647	98	332519	39.931	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	79.860%#
62) 4-Bromofluorobenzene	11.085	95	115452	34.346	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	68.700%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062123\
Data File : VX036263.D
Acq On : 21 Jun 2023 12:14
Operator : JC/MD
Sample : 03296-12
Misc : 15.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOTAL-VOC-R-01-GRAB-7

Quant Time: Jun 22 05:00:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

