

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062123\
 Data File : VX036264.D
 Acq On : 21 Jun 2023 12:38
 Operator : JC/MD
 Sample : 03331-05
 Misc : 20.53g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOTAL-VOC-SB8-GRAB-6

Quant Time: Jun 22 05:01:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	172121	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	306628	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	267541	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	118760	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	118102	37.547	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	75.100%#
35) Dibromofluoromethane	5.385	113	73913	35.547	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	71.100%#
50) Toluene-d8	8.647	98	305574	40.176	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	80.360%#
62) 4-Bromofluorobenzene	11.085	95	107194	34.914	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	69.820%#
Target Compounds						
					Qvalue	
18) Methyl Acetate	2.715	43	104640	21.113	ug/l	100
39) Methylcyclohexane	7.366	83	4460	1.352	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062123\
Data File : VX036264.D
Acq On : 21 Jun 2023 12:38
Operator : JC/MD
Sample : 03331-05
Misc : 20.53g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOTAL-VOC-SB8-GRAB-6

Quant Time: Jun 22 05:01:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

