

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062123\
 Data File : VX036274.D
 Acq On : 21 Jun 2023 16:32
 Operator : JC/MD
 Sample : 03296-05
 Misc : 13.76g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOTAL-VOC-R-02-GRAB-6

Quant Time: Jun 22 05:02:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	191253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	342107	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	288849	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130808	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	159506	45.638	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.280%
35) Dibromofluoromethane	5.385	113	103491	44.610	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.220%
50) Toluene-d8	8.647	98	423784	49.940	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.880%
62) 4-Bromofluorobenzene	11.085	95	146812	42.859	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062123\
Data File : VX036274.D
Acq On : 21 Jun 2023 16:32
Operator : JC/MD
Sample : 03296-05
Misc : 13.76g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOTAL-VOC-R-02-GRAB-6

Quant Time: Jun 22 05:02:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

