

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062123\
 Data File : VX036279.D
 Acq On : 21 Jun 2023 18:29
 Operator : JC/MD
 Sample : 03366-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB02

Quant Time: Jun 22 05:03:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	104203	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	204633	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	108361	56.905	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.800%
35) Dibromofluoromethane	5.385	113	73087	52.670	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.340%
50) Toluene-d8	8.647	98	284641	56.077	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	112.160%#
62) 4-Bromofluorobenzene	11.079	95	97774	47.719	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.440%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2233	2.822	ug/l #	72
20) Methylene Chloride	2.782	84	535	0.373	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062123\
Data File : VX036279.D
Acq On : 21 Jun 2023 18:29
Operator : JC/MD
Sample : 03366-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB02

Quant Time: Jun 22 05:03:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

