

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062123\
 Data File : VX036270.D
 Acq On : 21 Jun 2023 14:58
 Operator : JC/MD
 Sample : 03331-24
 Misc : 12.63g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOTAL-VOC-SB30-35-GRAB-5

Quant Time: Jun 22 05:02:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.544	168	175234	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	309492	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	259329	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	113081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	157198	49.089	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.180%
35) Dibromofluoromethane	5.385	113	102445	48.813	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.620%
50) Toluene-d8	8.647	98	396819	51.690	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.380%
62) 4-Bromofluorobenzene	11.085	95	143892	46.434	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.860%
Target Compounds						
18) Methyl Acetate	2.715	43	80657	15.985	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062123\
Data File : VX036270.D
Acq On : 21 Jun 2023 14:58
Operator : JC/MD
Sample : 03331-24
Misc : 12.63g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOTAL-VOC-SB30-35-GRAB-5


Quant Time: Jun 22 05:02:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.005 min
Lab File: VX036270.D
Acq: 21 Jun 2023 14:58

Instrument :
MSVOA_X
ClientSampleId :
TOTAL-VOC-SB30-35-GRAB-5







Tgt Ion:168 Resp: 175234
Ion Ratio Lower Upper
168 100
99 62.1 56.6 85.0



#18
Methyl Acetate
Concen: 15.985 ug/l
RT: 2.715 min Scan# 267
Delta R.T. 0.006 min
Lab File: VX036270.D
Acq: 21 Jun 2023 14:58

Tgt Ion: 43 Resp: 80657
Ion Ratio Lower Upper
43 100
74 21.8 17.8 26.8





#33

1,2-Dichloroethane-d4

Concen: 49.089 ug/l

RT: 5.952 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX036270.D

Acq: 21 Jun 2023 14:58

Instrument :

MSVOA_X

ClientSampleId :





TOTAL-VOC-SB30-35-GRAB-5

Tgt Ion: 65 Resp: 157198

Ion Ratio Lower Upper

65 100

67 52.0 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.006 min

Lab File: VX036270.D

Acq: 21 Jun 2023 14:58




Tgt Ion: 114 Resp: 309492


Ion Ratio Lower Upper

114 100

63 23.1 0.0 45.0


88 16.8 0.0 36.6








#35
Dibromofluoromethane
Concen: 48.813 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX036270.D
Acq: 21 Jun 2023 14:58

Instrument :
MSVOA_X
ClientSampleId :
TOTAL-VOC-SB30-35-GRAB-5







Tgt Ion:113 Resp: 102445
Ion Ratio Lower Upper
113 100
111 103.6 82.9 124.3
192 17.2 13.3 19.9



#50
Toluene-d8
Concen: 51.690 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.001 min
Lab File: VX036270.D
Acq: 21 Jun 2023 14:58

Tgt Ion: 98 Resp: 396819
Ion Ratio Lower Upper
98 100
100 66.6 53.1 79.7





#62

4-Bromofluorobenzene

Concen: 46.434 ug/l

RT: 11.085 min Scan# 1

Delta R.T. 0.006 min

Lab File: VX036270.D

Acq: 21 Jun 2023 14:58

Instrument :

MSVOA_X

ClientSampleId :

TOTAL-VOC-SB30-35-GRAB-5





Tgt Ion: 95 Resp: 143892

Ion Ratio Lower Upper

95 100

174 71.2 0.0 135.6

176 67.9 0.0 131.0



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VX036270.D

Acq: 21 Jun 2023 14:58




Tgt Ion: 117 Resp: 259329


Ion Ratio Lower Upper

117 100

82 57.5 46.2 69.4

119 31.8 25.5 38.3





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX036270.D


Acq: 21 Jun 2023 14:58

Instrument :

MSVOA_X

ClientSampleId :

TOTAL-VOC-SB30-35-GRAB-5



Tgt Ion:152 Resp: 113081

Ion Ratio Lower Upper

152 100

115 63.1 43.3 129.9

150 159.9 0.0 345.0

