

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062124\
 Data File : VX041968.D
 Acq On : 21 Jun 2024 10:01
 Operator : JC/MD
 Sample : VX0621MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0621MBL01

Quant Time: Jun 22 04:38:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	150834	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	232586	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	234125	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119351	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76997	44.269	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.540%
35) Dibromofluoromethane	5.385	113	72461	43.907	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.820%
50) Toluene-d8	8.647	98	275853	47.580	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.160%
62) 4-Bromofluorobenzene	11.079	95	105687	49.452	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062124\
Data File : VX041968.D
Acq On : 21 Jun 2024 10:01
Operator : JC/MD
Sample : VX0621MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0621MBL01

Quant Time: Jun 22 04:38:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

