

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062124\
 Data File : VX041978.D
 Acq On : 21 Jun 2024 14:00
 Operator : JC/MD
 Sample : P3003-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-189-TB-20240618

Quant Time: Jun 22 04:41:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	133321	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	203177	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	203608	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99580	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69118	44.959	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.920%
35) Dibromofluoromethane	5.385	113	65105	45.160	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.320%
50) Toluene-d8	8.653	98	240991	47.583	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.160%
62) 4-Bromofluorobenzene	11.079	95	88418	47.360	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.720%
Target Compounds						
29) Tetrahydrofuran	5.025	42	10580	15.599	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062124\
Data File : VX041978.D
Acq On : 21 Jun 2024 14:00
Operator : JC/MD
Sample : P3003-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-189-TB-20240618

Quant Time: Jun 22 04:41:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

