

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062124\
 Data File : VX041982.D
 Acq On : 21 Jun 2024 15:59
 Operator : JC/MD
 Sample : P3003-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-189-GW-583-585

Quant Time: Jun 22 04:43:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126262	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	194844	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193710	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96859	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64149	44.060	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.120%
35) Dibromofluoromethane	5.391	113	61419	44.425	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.860%
50) Toluene-d8	8.653	98	230861	47.533	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.060%
62) 4-Bromofluorobenzene	11.079	95	85361	47.678	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.360%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	26240	24.035	ug/l	89
12) 1,1-Dichloroethene	2.306	96	4900	4.499	ug/l	79
16) Acetone	2.386	43	2725	4.262	ug/l	97
27) cis-1,2-Dichloroethene	4.501	96	4150	2.998	ug/l	83
38) Carbon Tetrachloride	5.672	117	1847	1.139	ug/l #	81
44) Trichloroethene	7.123	130	469842	354.005	ug/l	89
55) 1,1,2-Trichloroethane	9.159	97	1105	0.841	ug/l	95
64) Tetrachloroethene	9.281	164	472	0.335	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062124\
Data File : VX041982.D
Acq On : 21 Jun 2024 15:59
Operator : JC/MD
Sample : P3003-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-189-GW-583-585

Quant Time: Jun 22 04:43:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

