

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062124\
 Data File : VX041989.D
 Acq On : 21 Jun 2024 18:45
 Operator : JC/MD
 Sample : P3005-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Jun 22 04:45:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	139336	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	217042	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216419	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107846	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72828	45.327	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.660%
35) Dibromofluoromethane	5.391	113	68095	44.217	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.440%
50) Toluene-d8	8.647	98	255041	47.141	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.280%
62) 4-Bromofluorobenzene	11.079	95	95870	48.071	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.140%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062124\
Data File : VX041989.D
Acq On : 21 Jun 2024 18:45
Operator : JC/MD
Sample : P3005-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Jun 22 04:45:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

