

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062218\
 Data File : VX002831.D
 Acq On : 22 Jun 2018 13:22
 Operator : JC/MD
 Sample : J3655-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPDES-1

Quant Time: Jun 22 13:47:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X061818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jun 18 15:18:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.02	128	42427	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	226518	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	208105	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.07	65	99348	30.66	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.20%
60) 4-Bromofluorobenzene	11.14	95	91960	26.76	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	89.20%
63) Toluene-d8	8.72	98	268114	29.49	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.30%

Target Compounds					Ovalue
25) Chloroform	5.21	83	19573	4.46 ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062218\
Data File : VX002831.D
Acq On : 22 Jun 2018 13:22
Operator : JC/MD
Sample : J3655-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPDES-1

Quant Time: Jun 22 13:47:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X061818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jun 18 15:18:02 2018
Response via : Initial Calibration

