

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062218\
 Data File : VX002842.D
 Acq On : 22 Jun 2018 20:26
 Operator : JC/MD
 Sample : J3604-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-10

Quant Time: Jun 25 11:02:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:37:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	221751	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	312833	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	294659	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	163120	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	153186	45.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.92%	
35) Dibromofluoromethane	5.51	113	118645	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
50) Toluene-d8	8.72	98	441301	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
62) 4-Bromofluorobenzene	11.14	95	154088	42.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.78%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	15598	5.70	ug/l	83
30) Chloroform	5.21	83	6224	1.34	ug/l	96
44) Trichloroethene	7.22	130	9988	3.47	ug/l	96
64) Tetrachloroethene	9.34	164	571157	174.62	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062218\
Data File : VX002842.D
Acq On : 22 Jun 2018 20:26
Operator : JC/MD
Sample : J3604-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-10

Quant Time: Jun 25 11:02:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

