

Data File : VX002843.D
Acq On : 22 Jun 2018 20:52
Operator : JC/MD
Sample : J3592-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0626

Quant Time: Jul 09 05:03:51 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	238107	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	336781	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	315960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	177843	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	163736	45.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.50%	
35) Dibromofluoromethane	5.51	113	129610	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
50) Toluene-d8	8.72	98	480284	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
62) 4-Bromofluorobenzene	11.14	95	167157	43.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.44%	
Target Compounds						
2) Dichlorodifluoromethane	1.20	85	3774	1.493	ug/l	93
21) trans-1,2-Dichloroethene	3.17	96	4399	1.524	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	27296	9.295	ug/l	98
44) Trichloroethene	7.22	130	6471	2.090	ug/l	97
64) Tetrachloroethene	9.34	164	3276	0.934	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX002843.D
Acq On : 22 Jun 2018 20:52
Operator : JC/MD
Sample : J3592-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0626

Quant Time: Jul 09 05:03:51 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

