

Data File : VX002844.D
Acq On : 22 Jun 2018 21:19
Operator : JC/MD
Sample : J3592-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0627

Quant Time: Jul 09 05:04:56 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	228725	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	316820	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	295193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	169284	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.07	65	157053	45.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.38%	
35) Dibromofluoromethane	5.51	113	123590	48.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.76%	
50) Toluene-d8	8.72	98	452168	47.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.20%	
62) 4-Bromofluorobenzene	11.14	95	163175	44.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.70%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.20	85	3875	1.596	ug/l	95
4) Vinyl Chloride	1.40	62	310042	103.440	ug/l	91
12) 1,1-Dichloroethene	2.37	96	2948	1.153	ug/l	89
16) Acetone	2.45	43	20082	13.096	ug/l	97
21) trans-1,2-Dichloroethene	3.17	96	26291	9.483	ug/l	99
25) 2-Butanone	4.71	43	4811	2.389	ug/l #	86
27) cis-1,2-Dichloroethene	4.60	96	433096	153.523	ug/l	83
44) Trichloroethene	7.22	130	62685	21.525	ug/l	99
59) 2-Hexanone	9.50	43	6081	1.992	ug/l	95
64) Tetrachloroethene	9.34	164	1096	0.334	ug/l #	87
78) n-propylbenzene	11.37	91	5274	0.406	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX002844.D
Acq On : 22 Jun 2018 21:19
Operator : JC/MD
Sample : J3592-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0627

Quant Time: Jul 09 05:04:56 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061518W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:37:33 2018
Response via : Initial Calibration

