

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX062220\
 Data File : VX016926.D
 Acq On : 22 Jun 2020 17:44
 Operator : JC/SP
 Sample : L3032-02DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF015MW047-200616DL

Quant Time: Jun 23 08:12:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 19 12:07:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	307390	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	567135	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	506964	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	309115	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	168412	45.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.10%	
35) Dibromofluoromethane	5.46	113	148661	40.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.96%	
50) Toluene-d8	8.70	98	702674	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
62) 4-Bromofluorobenzene	11.12	95	257377	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.01	59	8255	13.209	ug/l	99
16) Acetone	2.42	43	2955	2.202	ug/l #	85
22) Diisopropyl ether	3.82	45	10575	1.270	ug/l #	81
40) Benzene	6.11	78	10820	0.689	ug/l	100
65) Chlorobenzene	10.12	112	572539	55.770	ug/l	98
88) 1,4-Dichlorobenzene	12.08	146	34946	3.485	ug/l	89
91) 1,2-Dichlorobenzene	12.38	146	144750	15.120	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX062220\
Data File : VX016926.D
Acq On : 22 Jun 2020 17:44
Operator : JC/SP
Sample : L3032-02DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF015MW047-200616DL

Quant Time: Jun 23 08:12:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 19 12:07:48 2020
Response via : Initial Calibration

