


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX062220\
 Data File : VX016921.D
 Acq On : 22 Jun 2020 14:08
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Title : SW846 8260
 Last Update : Fri Jun 19 12:07:48 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1637

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	21459	PASS
75	95	30	60	46.8	62604	PASS
95	95	100	100	100.0	133826	PASS
96	95	5	9	6.6	8877	PASS
173	174	0.00	2	1.0	1091	PASS
174	95	50	100	85.6	114533	PASS
175	174	5	9	7.4	8459	PASS
176	174	95	101	99.0	113394	PASS
177	176	5	9	6.7	7557	PASS