

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029686.D
 Acq On : 22 Jun 2022 09:56
 Operator : JC/MD
 Sample : VX0622MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0622MBL01

Quant Time: Jun 23 01:07:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	276100	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	463181	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	403772	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	176568	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	146705	40.140	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	80.280%
35) Dibromofluoromethane	5.385	113	135541	44.904	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.800%
50) Toluene-d8	8.646	98	520491	47.201	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.400%
62) 4-Bromofluorobenzene	11.079	95	188016	45.213	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029686.D
Acq On : 22 Jun 2022 09:56
Operator : JC/MD
Sample : VX0622MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0622MBL01

Quant Time: Jun 23 01:07:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

