

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029687.D
 Acq On : 22 Jun 2022 10:19
 Operator : JC/MD
 Sample : VX0622WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0622WBL01

Quant Time: Jun 23 01:07:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	273855	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	459647	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	401852	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	177510	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	150059	41.395	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.780%
35) Dibromofluoromethane	5.391	113	137602	45.937	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.880%
50) Toluene-d8	8.653	98	520871	47.599	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.200%
62) 4-Bromofluorobenzene	11.079	95	186983	45.310	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029687.D
Acq On : 22 Jun 2022 10:19
Operator : JC/MD
Sample : VX0622WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0622WBL01

Quant Time: Jun 23 01:07:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

