

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029691.D
 Acq On : 22 Jun 2022 12:07
 Operator : JC/MD
 Sample : N3286-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6S

Quant Time: Jun 23 01:08:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	262995	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	446820	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	387665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	175269	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	147791	42.452	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.900%
35) Dibromofluoromethane	5.391	113	132109	45.369	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.740%
50) Toluene-d8	8.653	98	507842	47.740	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.480%
62) 4-Bromofluorobenzene	11.079	95	184025	45.874	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.740%
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.520	76	4655	0.674	ug/l	98
31) Cyclohexane	5.483	56	4121	0.882	ug/l #	81
40) Benzene	6.050	78	7646	0.655	ug/l	97
67) Ethyl Benzene	10.195	91	7068	0.517	ug/l	98
68) m/p-Xylenes	10.305	106	4228	0.792	ug/l	96
73) Isopropylbenzene	10.964	105	5056	0.385	ug/l	95
83) tert-Butylbenzene	11.713	119	2623	0.243	ug/l	82
84) 1,2,4-Trimethylbenzene	11.756	105	5312	0.481	ug/l	97
85) sec-Butylbenzene	11.890	105	4309	0.322	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029691.D
Acq On : 22 Jun 2022 12:07
Operator : JC/MD
Sample : N3286-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6S

Quant Time: Jun 23 01:08:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

