

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029694.D
 Acq On : 22 Jun 2022 13:17
 Operator : JC/MD
 Sample : N3354-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-62

Quant Time: Jun 23 01:09:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	259012	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	434319	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	373426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	167430	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	143199	41.766	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	83.540%		
35) Dibromofluoromethane	5.391	113	128968	45.565	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.140%		
50) Toluene-d8	8.653	98	490945	47.480	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	94.960%		
62) 4-Bromofluorobenzene	11.079	95	175466	44.999	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.000%		
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.520	76	2528	0.372	ug/l #	91
39) Methylcyclohexane	7.379	83	1707	0.357	ug/l #	70
64) Tetrachloroethene	9.275	164	589	0.218	ug/l #	79
68) m/p-Xylenes	10.311	106	3284	0.639	ug/l	98
69) o-Xylene	10.640	106	1136	0.224	ug/l	97
73) Isopropylbenzene	10.963	105	16101	1.283	ug/l	95
78) n-propylbenzene	11.305	91	17559	1.242	ug/l	97
85) sec-Butylbenzene	11.890	105	9744	0.763	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029694.D
Acq On : 22 Jun 2022 13:17
Operator : JC/MD
Sample : N3354-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-62

Quant Time: Jun 23 01:09:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

