

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029699.D
 Acq On : 22 Jun 2022 15:14
 Operator : JC/MD
 Sample : N3293-15 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40807

Quant Time: Jun 22 15:46:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	257679	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	431625	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	373177	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	164218	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140916	41.313	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.620%
35) Dibromofluoromethane	5.391	113	124301	44.191	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.380%
50) Toluene-d8	8.653	98	488436	47.532	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.060%
62) 4-Bromofluorobenzene	11.085	95	171900	44.360	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.720%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	29943	22.024	ug/l	100
20) Methylene Chloride	2.788	84	1877	0.620	ug/l	94
25) 2-Butanone	4.562	43	618742	287.916	ug/l	99
40) Benzene	6.043	78	74640	6.614	ug/l	96
52) Toluene	8.720	92	384695	53.236	ug/l	97
67) Ethyl Benzene	10.195	91	122895	9.330	ug/l	99
68) m/p-Xylenes	10.305	106	188499	36.695	ug/l	97
69) o-Xylene	10.646	106	115560	22.769	ug/l	100
70) Styrene	10.659	104	65897	7.930	ug/l #	81
73) Isopropylbenzene	10.963	105	6707	0.545	ug/l	99
78) n-propylbenzene	11.305	91	22208	1.601	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	41704	4.022	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	223353	21.597	ug/l	99
95) Naphthalene	13.780	128	251815	21.861	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029699.D
Acq On : 22 Jun 2022 15:14
Operator : JC/MD
Sample : N3293-15 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40807

Quant Time: Jun 22 15:46:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

