

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029700.D
 Acq On : 22 Jun 2022 15:37
 Operator : JC/MD
 Sample : N3293-18 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 52322-D

Quant Time: Jun 22 16:10:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	253877	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	431762	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	374508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	169461	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	142250	42.328	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.660%
35) Dibromofluoromethane	5.391	113	125960	44.766	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.540%
50) Toluene-d8	8.652	98	484760	47.160	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.320%
62) 4-Bromofluorobenzene	11.085	95	175187	45.194	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.380%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	332405	248.153	ug/l	99
52) Toluene	8.720	92	20903	2.892	ug/l	100
68) m/p-Xylenes	10.305	106	3858	0.748	ug/l	93
69) o-Xylene	10.640	106	2077	0.408	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	4083	0.382	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	14523	1.361	ug/l	93
95) Naphthalene	13.780	128	313964	26.413	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029700.D
Acq On : 22 Jun 2022 15:37
Operator : JC/MD
Sample : N3293-18 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
52322-D

Quant Time: Jun 22 16:10:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

