

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029703.D
 Acq On : 22 Jun 2022 16:47
 Operator : JC/MD
 Sample : N3293-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40806

Quant Time: Jun 22 17:16:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	253508	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	430297	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	369676	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	139046	41.435	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.880%
35) Dibromofluoromethane	5.391	113	127353	45.415	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.840%
50) Toluene-d8	8.653	98	482955	47.144	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.280%
62) 4-Bromofluorobenzene	11.085	95	171237	44.325	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.640%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	10877	8.132	ug/l	91
20) Methylene Chloride	2.794	84	3649	1.225	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029703.D
Acq On : 22 Jun 2022 16:47
Operator : JC/MD
Sample : N3293-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40806

Quant Time: Jun 22 17:16:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

