

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029708.D
 Acq On : 22 Jun 2022 18:43
 Operator : JC/MD
 Sample : N3202-09RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-31RE

Quant Time: Jun 23 01:10:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270669	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	455249	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	395159	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	171889	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149358	41.686	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.380%
35) Dibromofluoromethane	5.391	113	133837	45.112	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.220%
50) Toluene-d8	8.653	98	514173	47.441	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.880%
62) 4-Bromofluorobenzene	11.079	95	183743	44.955	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.920%
Target Compounds						
					Qvalue	
68) m/p-Xylenes	10.305	106	1285	0.236	ug/l	92
84) 1,2,4-Trimethylbenzene	11.756	105	6936	0.641	ug/l	99
95) Naphthalene	13.780	128	3756	0.312	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029708.D
Acq On : 22 Jun 2022 18:43
Operator : JC/MD
Sample : N3202-09RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-31RE

Quant Time: Jun 23 01:10:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

