

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029709.D
 Acq On : 22 Jun 2022 19:07
 Operator : JC/MD
 Sample : N3202-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22

Quant Time: Jun 23 01:10:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	271333	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	460907	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	405212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	175816	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	148384	41.313	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.620%
35) Dibromofluoromethane	5.391	113	136150	45.328	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.660%
50) Toluene-d8	8.653	98	520833	47.465	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.940%
62) 4-Bromofluorobenzene	11.085	95	189199	45.722	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.440%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.965	59	467690	610.886	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	1823737	191.102	ug/l	100
22) Diisopropyl ether	3.757	45	10693	1.218	ug/l #	82
40) Benzene	6.037	78	14858	1.233	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	5283	0.477	ug/l	99
95) Naphthalene	13.780	128	3350	0.272	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029709.D
Acq On : 22 Jun 2022 19:07
Operator : JC/MD
Sample : N3202-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22

Quant Time: Jun 23 01:10:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

