

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029710.D
 Acq On : 22 Jun 2022 19:30
 Operator : JC/MD
 Sample : N3202-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-25

Quant Time: Jun 23 01:11:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	272712	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	454494	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	394186	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	177366	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	148405	41.110	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.220%
35) Dibromofluoromethane	5.391	113	134192	45.307	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.620%
50) Toluene-d8	8.653	98	514785	47.576	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.160%
62) 4-Bromofluorobenzene	11.079	95	186163	45.623	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.240%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.959	59	84991	102.580	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	328197	34.217	ug/l	99
22) Diisopropyl ether	3.770	45	23699	2.686	ug/l #	83
84) 1,2,4-Trimethylbenzene	11.756	105	3288	0.294	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029710.D
Acq On : 22 Jun 2022 19:30
Operator : JC/MD
Sample : N3202-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-25

Quant Time: Jun 23 01:11:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

