

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029714.D
 Acq On : 23 Jun 2022 02:16
 Operator : JC/MD
 Sample : VX0622WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0622WBL02

Quant Time: Jun 23 04:12:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	285307	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	476955	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	401911	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	169518	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149055	39.467	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	78.940%
35) Dibromofluoromethane	5.385	113	139494	44.879	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.760%
50) Toluene-d8	8.653	98	531545	46.811	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.620%
62) 4-Bromofluorobenzene	11.079	95	184424	43.068	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029714.D
Acq On : 23 Jun 2022 02:16
Operator : JC/MD
Sample : VX0622WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0622WBL02

Quant Time: Jun 23 04:12:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

