

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029722.D
 Acq On : 23 Jun 2022 05:21
 Operator : JC/MD
 Sample : N3293-01
 Misc : 1.05g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 30873

Quant Time: Jun 23 07:20:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	302725	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	499324	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	422348	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	192345	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	158002	39.429	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	78.860%
35) Dibromofluoromethane	5.391	113	138738	42.636	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.280%
50) Toluene-d8	8.653	98	558000	46.940	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.880%
62) 4-Bromofluorobenzene	11.085	95	196572	43.849	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029722.D
Acq On : 23 Jun 2022 05:21
Operator : JC/MD
Sample : N3293-01
Misc : 1.05g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30873

Quant Time: Jun 23 07:20:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

