

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029723.D
 Acq On : 23 Jun 2022 05:44
 Operator : JC/MD
 Sample : N3334-05 10X
 Misc : 1.79g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-14A-1A-7-GR

Quant Time: Jun 23 07:20:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	291666	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	485098	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	425344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	193185	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156962	40.655	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.300%
35) Dibromofluoromethane	5.385	113	138629	43.852	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.700%
50) Toluene-d8	8.653	98	550155	47.637	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.280%
62) 4-Bromofluorobenzene	11.085	95	203776	46.789	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.580%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.470	56	8929	1.723	ug/l	93
39) Methylcyclohexane	7.379	83	51858	9.721	ug/l	94
68) m/p-Xylenes	10.305	106	6745	1.152	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	17696	1.451	ug/l	100
84) 1,2,4-Trimethylbenzene	11.756	105	64734	5.321	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029723.D
Acq On : 23 Jun 2022 05:44
Operator : JC/MD
Sample : N3334-05 10X
Misc : 1.79g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-14A-1A-7-GR

Quant Time: Jun 23 07:20:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

