

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029728.D
 Acq On : 23 Jun 2022 07:40
 Operator : JC/MD
 Sample : N3422-11 40X
 Misc : 5.76g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-14A-1A-13-GR

Quant Time: Jun 23 09:09:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	288577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	485260	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	418352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	194283	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	155321	40.660	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.320%
35) Dibromofluoromethane	5.391	113	139131	43.996	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.000%
50) Toluene-d8	8.653	98	538464	46.609	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.220%
62) 4-Bromofluorobenzene	11.079	95	201425	46.234	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.460%
Target Compounds						
					Qvalue	
39) Methylcyclohexane	7.385	83	8725	1.635	ug/l	92
86) p-Isopropyltoluene	12.012	119	12660	1.020	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029728.D
Acq On : 23 Jun 2022 07:40
Operator : JC/MD
Sample : N3422-11 40X
Misc : 5.76g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-14A-1A-13-GR

Quant Time: Jun 23 09:09:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

