

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX062222\
 Data File : VX029729.D
 Acq On : 23 Jun 2022 08:03
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 23 09:09:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	50.000	51.345	-2.7	86	0.00
3 P	Chloromethane	50.000	43.761	12.5	81	0.00
4 C	Vinyl Chloride	50.000	49.384	1.2#	87	0.00
5 T	Bromomethane	50.000	44.803	10.4	88	0.00
6 T	Chloroethane	50.000	38.767	22.5	73	0.00
7 T	Trichlorofluoromethane	50.000	50.630	-1.3	89	0.00
8 T	Diethyl Ether	50.000	51.464	-2.9	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.610	-3.2	93	0.00
10 T	Methyl Iodide	50.000	42.628	14.7	77	0.00
11 T	Tert butyl alcohol	250.000	244.884	2.0	93	0.02
12 CM	1,1-Dichloroethene	50.000	53.969	-7.9#	96	0.00
13 T	Acrolein	250.000	205.451	17.8	79	0.00
14 T	Allyl chloride	50.000	48.916	2.2	86	0.00
15 T	Acrylonitrile	250.000	259.658	-3.9	91	0.00
16 T	Acetone	250.000	229.807	8.1	87	0.00
17 T	Carbon Disulfide	50.000	48.314	3.4	86	0.00
18 T	Methyl Acetate	50.000	51.014	-2.0	92	0.00
19 T	Methyl tert-butyl Ether	50.000	53.454	-6.9	95	0.00
20 T	Methylene Chloride	50.000	52.344	-4.7	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	52.782	-5.6	93	0.00
22 T	Diisopropyl ether	50.000	51.745	-3.5	92	0.00
23 T	Vinyl Acetate	250.000	257.743	-3.1	88	0.00
24 P	1,1-Dichloroethane	50.000	52.206	-4.4	93	0.00
25 T	2-Butanone	250.000	243.967	2.4	87	0.00
26 T	2,2-Dichloropropane	50.000	35.981	28.0#	62	0.00
27 T	cis-1,2-Dichloroethene	50.000	54.129	-8.3	96	0.00
28 T	Bromochloromethane	50.000	49.323	1.4	89	0.00
29 T	Tetrahydrofuran	250.000	245.608	1.8	86	0.00
30 C	Chloroform	50.000	51.848	-3.7#	92	0.00
31 T	Cyclohexane	50.000	50.524	-1.0	90	0.00
32 T	1,1,1-Trichloroethane	50.000	51.602	-3.2	91	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.854	6.3	84	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	92	0.00
35 S	Dibromofluoromethane	50.000	50.140	-0.3	90	0.00
36 T	1,1-Dichloropropene	50.000	50.164	-0.3	90	0.00
37 T	Ethyl Acetate	50.000	49.051	1.9	85	0.00
38 T	Carbon Tetrachloride	50.000	50.007	-0.0	88	0.00
39 T	Methylcyclohexane	50.000	49.234	1.5	87	0.00
40 TM	Benzene	50.000	52.181	-4.4	94	0.00
41 T	Methacrylonitrile	50.000	51.794	-3.6	88	0.00
42 TM	1,2-Dichloroethane	50.000	48.911	2.2	90	0.00
43 T	Isopropyl Acetate	50.000	49.983	0.0	87	0.00
44 TM	Trichloroethene	50.000	52.452	-4.9	95	0.00
45 C	1,2-Dichloropropane	50.000	52.828	-5.7#	96	0.00
46 T	Dibromomethane	50.000	51.702	-3.4	92	0.00
47 T	Bromodichloromethane	50.000	50.314	-0.6	88	0.00
48 T	Methyl methacrylate	50.000	49.320	1.4	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029729.D
 Acq On : 23 Jun 2022 08:03
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 23 09:09:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	909.366	9.1	83	0.00
50 S	Toluene-d8	50.000	50.646	-1.3	89	0.00
51 T	4-Methyl-2-Pentanone	250.000	244.225	2.3	87	0.00
52 CM	Toluene	50.000	53.291	-6.6#	94	0.00
53 T	t-1,3-Dichloropropene	50.000	50.582	-1.2	84	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.924	-1.8	86	0.00
55 T	1,1,2-Trichloroethane	50.000	53.136	-6.3	94	0.00
56 T	Ethyl methacrylate	50.000	54.150	-8.3	92	0.00
57 T	1,3-Dichloropropane	50.000	51.624	-3.2	93	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	249.191	0.3	86	0.00
59 T	2-Hexanone	250.000	246.547	1.4	86	0.00
60 T	Dibromochloromethane	50.000	52.827	-5.7	90	0.00
61 T	1,2-Dibromoethane	50.000	54.098	-8.2	94	0.00
62 S	4-Bromofluorobenzene	50.000	51.764	-3.5	89	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	92	0.00
64 T	Tetrachloroethene	50.000	52.231	-4.5	95	0.00
65 PM	Chlorobenzene	50.000	53.333	-6.7	95	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.796	-5.6	92	0.00
67 C	Ethyl Benzene	50.000	52.149	-4.3#	92	0.00
68 T	m/p-Xylenes	100.000	105.376	-5.4	94	0.00
69 T	o-Xylene	50.000	52.303	-4.6	93	0.00
70 T	Styrene	50.000	53.827	-7.7	93	0.00
71 P	Bromoform	50.000	51.854	-3.7	87	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	93	0.00
73 T	Isopropylbenzene	50.000	50.353	-0.7	93	0.00
74 T	N-amyl acetate	50.000	51.934	-3.9	92	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.283	1.4	93	0.00
76 T	1,2,3-Trichloropropane	50.000	48.715	2.6	89	0.00
77 T	Bromobenzene	50.000	51.600	-3.2	95	0.00
78 T	n-propylbenzene	50.000	50.195	-0.4	91	0.00
79 T	2-Chlorotoluene	50.000	49.560	0.9	92	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.504	-1.0	92	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	43.152	13.7	73	0.00
82 T	4-Chlorotoluene	50.000	49.484	1.0	92	0.00
83 T	tert-Butylbenzene	50.000	50.286	-0.6	93	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.442	-0.9	93	0.00
85 T	sec-Butylbenzene	50.000	50.282	-0.6	91	0.00
86 T	p-Isopropyltoluene	50.000	50.505	-1.0	91	0.00
87 T	1,3-Dichlorobenzene	50.000	51.735	-3.5	95	0.00
88 T	1,4-Dichlorobenzene	50.000	52.051	-4.1	95	0.00
89 T	n-Butylbenzene	50.000	49.985	0.0	89	0.00
90 T	Hexachloroethane	50.000	50.823	-1.6	86	0.00
91 T	1,2-Dichlorobenzene	50.000	52.018	-4.0	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.575	4.8	88	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.742	-3.5	92	0.00
94 T	Hexachlorobutadiene	50.000	47.828	4.3	85	0.00
95 T	Naphthalene	50.000	55.430	-10.9	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
Data File : VX029729.D
Acq On : 23 Jun 2022 08:03
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 23 09:09:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.686	-5.4	95	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6