



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062222\
 Data File : VX029684.D
 Acq On : 22 Jun 2022 08:14
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Title : SW846 8260
 Last Update : Mon Jun 20 01:31:01 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.8	9969	PASS
75	95	30	60	48.5	30627	PASS
95	95	100	100	100.0	63195	PASS
96	95	5	9	6.9	4334	PASS
173	174	0.00	2	1.0	457	PASS
174	95	50	100	71.0	44845	PASS
175	174	5	9	7.4	3338	PASS
176	174	95	101	96.1	43077	PASS
177	176	5	9	6.7	2886	PASS