


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX062320\
 Data File : VX016931.D
 Acq On : 23 Jun 2020 15:12
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062320W.M
 Title : SW846 8260
 Last Update : Wed Jun 24 01:00:35 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1637

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.4	18575	PASS
75	95	30	60	52.7	48106	PASS
95	95	100	100	100.0	91200	PASS
96	95	5	9	6.5	5913	PASS
173	174	0.00	2	0.6	404	PASS
174	95	50	100	71.3	65040	PASS
175	174	5	9	8.1	5249	PASS
176	174	95	101	95.0	61805	PASS
177	176	5	9	6.9	4250	PASS