

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062321\
 Data File : VX022847.D
 Acq On : 23 Jun 2021 18:08
 Operator : JC/MD
 Sample : VX0623WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0623WBL01

Quant Time: Jun 24 03:36:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	54467	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	109249	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	96745	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36049	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53049	53.67	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.34%	
35) Dibromofluoromethane	5.397	113	36879	53.16	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.32%	
50) Toluene-d8	8.653	98	146346	53.82	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.64%	
62) 4-Bromofluorobenzene	11.085	95	48884	48.17	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.34%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062321\
Data File : VX022847.D
Acq On : 23 Jun 2021 18:08
Operator : JC/MD
Sample : VX0623WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0623WBL01

Quant Time: Jun 24 03:36:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

