

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062321\
 Data File : VX022850.D
 Acq On : 23 Jun 2021 19:18
 Operator : JC/MD
 Sample : M2783-11 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE139D2-20210617

Quant Time: Jun 24 03:37:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	55790	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	115821	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105107	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36742	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54039	53.37	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.74%	
35) Dibromofluoromethane	5.397	113	38290	52.06	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.12%	
50) Toluene-d8	8.653	98	151316	52.49	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.98%	
62) 4-Bromofluorobenzene	11.085	95	51875	48.22	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.44%	
Target Compounds						Qvalue
44) Trichloroethene	7.135	130	30628	36.41	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062321\
Data File : VX022850.D
Acq On : 23 Jun 2021 19:18
Operator : JC/MD
Sample : M2783-11 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE139D2-20210617

Quant Time: Jun 24 03:37:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

