

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062321\
 Data File : VX022851.D
 Acq On : 23 Jun 2021 19:42
 Operator : JC/MD
 Sample : M2783-18 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20210618

Quant Time: Jun 24 03:37:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53899	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	110709	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	95945	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35295	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51556	52.71	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.42%	
35) Dibromofluoromethane	5.397	113	35626	50.68	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.36%	
50) Toluene-d8	8.653	98	145707	52.88	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.76%	
62) 4-Bromofluorobenzene	11.085	95	48307	46.98	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.96%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	986	1.41	ug/l	92
27) cis-1,2-Dichloroethene	4.501	96	490	0.57	ug/l	92
44) Trichloroethene	7.135	130	41480	51.58	ug/l	89
64) Tetrachloroethene	9.275	164	778	1.15	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062321\
Data File : VX022851.D
Acq On : 23 Jun 2021 19:42
Operator : JC/MD
Sample : M2783-18 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210618

Quant Time: Jun 24 03:37:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

