

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062321\
 Data File : VX022855.D
 Acq On : 23 Jun 2021 21:16
 Operator : JC/MD
 Sample : M2803-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW203D-20210618

Quant Time: Jun 24 03:38:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	50673	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	104258	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	91610	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	34353	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48344	52.57	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.14%			
35) Dibromofluoromethane	5.397	113	34086	51.48	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.96%			
50) Toluene-d8	8.653	98	133998	51.64	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 103.28%			
62) 4-Bromofluorobenzene	11.085	95	46455	47.97	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.94%			
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.319	96	1111	1.69	ug/l	# 72
16) Acetone	2.386	43	1597	3.92	ug/l	# 87
24) 1,1-Dichloroethane	3.617	63	4711	3.39	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	4328	5.35	ug/l	97
30) Chloroform	5.105	83	1076	0.77	ug/l	# 68
32) 1,1,1-Trichloroethane	5.391	97	1034	0.92	ug/l	# 49
44) Trichloroethene	7.129	130	19778	26.12	ug/l	92
64) Tetrachloroethene	9.275	164	517	0.80	ug/l	# 62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062321\
Data File : VX022855.D
Acq On : 23 Jun 2021 21:16
Operator : JC/MD
Sample : M2803-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW203D-20210618

Quant Time: Jun 24 03:38:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

