

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062321\
 Data File : VX022846.D
 Acq On : 23 Jun 2021 15:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 24 03:36:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	50.000	49.566	0.9	87	0.00
3 P	Chloromethane	50.000	47.662	4.7	91	0.00
4 C	Vinyl Chloride	50.000	47.260	5.5#	86	0.00
5 T	Bromomethane	50.000	50.395	-0.8	100	0.00
6 T	Chloroethane	50.000	49.425	1.2	90	0.00
7 T	Trichlorofluoromethane	50.000	48.619	2.8	90	0.00
8 T	Diethyl Ether	50.000	53.014	-6.0	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.106	-0.2	92	0.00
10 T	Methyl Iodide	50.000	41.052	17.9	80	0.00
11 T	Tert butyl alcohol	250.000	255.942	-2.4	98	0.00
12 CM	1,1-Dichloroethene	50.000	47.097	5.8#	88	0.00
13 T	Acrolein	250.000	239.721	4.1	94	0.00
14 T	Allyl chloride	50.000	49.716	0.6	92	0.00
15 T	Acrylonitrile	250.000	272.082	-8.8	98	0.00
16 T	Acetone	250.000	242.777	2.9	96	0.00
17 T	Carbon Disulfide	50.000	43.401	13.2	89	0.00
18 T	Methyl Acetate	50.000	52.488	-5.0	100	0.00
19 T	Methyl tert-butyl Ether	50.000	54.000	-8.0	98	0.00
20 T	Methylene Chloride	50.000	50.069	-0.1	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.183	3.6	91	0.00
22 T	Diisopropyl ether	50.000	53.357	-6.7	96	0.00
23 T	Vinyl Acetate	250.000	283.270	-13.3	99	0.00
24 P	1,1-Dichloroethane	50.000	50.677	-1.4	93	0.00
25 T	2-Butanone	250.000	265.478	-6.2	97	0.00
26 T	2,2-Dichloropropane	50.000	50.432	-0.9	90	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.929	-3.9	95	0.00
28 T	Bromochloromethane	50.000	50.792	-1.6	95	0.00
29 T	Tetrahydrofuran	250.000	274.422	-9.8	98	0.00
30 C	Chloroform	50.000	49.542	0.9#	92	0.00
31 T	Cyclohexane	50.000	48.393	3.2	90	0.00
32 T	1,1,1-Trichloroethane	50.000	49.873	0.3	90	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.988	0.0	99	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	91	0.00
35 S	Dibromofluoromethane	50.000	51.246	-2.5	95	0.00
36 T	1,1-Dichloropropene	50.000	49.729	0.5	92	0.00
37 T	Ethyl Acetate	50.000	56.007	-12.0	101	0.00
38 T	Carbon Tetrachloride	50.000	52.403	-4.8	89	0.00
39 T	Methylcyclohexane	50.000	52.049	-4.1	92	0.00
40 TM	Benzene	50.000	51.753	-3.5	93	0.00
41 T	Methacrylonitrile	50.000	54.762	-9.5	101	0.00
42 TM	1,2-Dichloroethane	50.000	54.740	-9.5	95	0.00
43 T	Isopropyl Acetate	50.000	56.580	-13.2	100	0.00
44 TM	Trichloroethene	50.000	50.537	-1.1	91	0.00
45 C	1,2-Dichloropropane	50.000	54.161	-8.3#	96	0.00
46 T	Dibromomethane	50.000	54.851	-9.7	100	0.00
47 T	Bromodichloromethane	50.000	49.284	1.4	96	0.00
48 T	Methyl methacrylate	50.000	57.290	-14.6	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062321\
 Data File : VX022846.D
 Acq On : 23 Jun 2021 15:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 24 03:36:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 23 19:12:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1114.869	-11.5	97	0.00
50 S	Toluene-d8	50.000	50.860	-1.7	95	0.00
51 T	4-Methyl-2-Pentanone	250.000	290.138	-16.1	99	0.00
52 CM	Toluene	50.000	53.679	-7.4#	94	0.00
53 T	t-1,3-Dichloropropene	50.000	48.656	2.7	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.949	0.1	95	0.00
55 T	1,1,2-Trichloroethane	50.000	55.758	-11.5	98	0.00
56 T	Ethyl methacrylate	50.000	51.165	-2.3	97	0.00
57 T	1,3-Dichloropropane	50.000	54.517	-9.0	96	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	273.041	-9.2	100	0.00
59 T	2-Hexanone	250.000	288.640	-15.5	98	0.00
60 T	Dibromochloromethane	50.000	49.116	1.8	96	0.00
61 T	1,2-Dibromoethane	50.000	55.314	-10.6	98	0.00
62 S	4-Bromofluorobenzene	50.000	50.475	-1.0	95	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	96	0.00
64 T	Tetrachloroethene	50.000	48.054	3.9	89	0.00
65 PM	Chlorobenzene	50.000	50.115	-0.2	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.510	-5.0	92	0.00
67 C	Ethyl Benzene	50.000	50.573	-1.1#	92	0.00
68 T	m/p-Xylenes	100.000	100.298	-0.3	93	0.00
69 T	o-Xylene	50.000	50.105	-0.2	92	0.00
70 T	Styrene	50.000	52.258	-4.5	95	0.00
71 P	Bromoform	50.000	45.763	8.5	99	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	93	0.00
73 T	Isopropylbenzene	50.000	51.602	-3.2	92	0.00
74 T	N-ethyl acetate	50.000	50.701	-1.4	97	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	55.610	-11.2	99	0.00
76 T	1,2,3-Trichloropropane	50.000	56.217	-12.4	100	0.00
77 T	Bromobenzene	50.000	51.633	-3.3	95	0.00
78 T	n-propylbenzene	50.000	52.907	-5.8	93	0.00
79 T	2-Chlorotoluene	50.000	51.458	-2.9	92	0.00
80 T	1,3,5-Trimethylbenzene	50.000	54.540	-9.1	96	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.262	1.5	100	0.00
82 T	4-Chlorotoluene	50.000	51.526	-3.1	95	0.00
83 T	tert-Butylbenzene	50.000	51.361	-2.7	92	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.776	-7.6	93	0.00
85 T	sec-Butylbenzene	50.000	53.001	-6.0	93	0.00
86 T	p-Isopropyltoluene	50.000	52.641	-5.3	92	0.00
87 T	1,3-Dichlorobenzene	50.000	52.092	-4.2	94	0.00
88 T	1,4-Dichlorobenzene	50.000	50.604	-1.2	95	0.00
89 T	n-Butylbenzene	50.000	54.691	-9.4	94	0.00
90 T	Hexachloroethane	50.000	46.869	6.3	95	0.00
91 T	1,2-Dichlorobenzene	50.000	51.959	-3.9	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.015	2.0	101	0.00
93 T	1,2,4-Trichlorobenzene	50.000	54.732	-9.5	97	0.00
94 T	Hexachlorobutadiene	50.000	52.787	-5.6	102	0.00
95 T	Naphthalene	50.000	50.959	-1.9	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062321\
Data File : VX022846.D
Acq On : 23 Jun 2021 15:28
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 24 03:36:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062321W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 23 19:12:26 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	53.785	-7.6	92	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6