

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062322\
 Data File : VX029739.D
 Acq On : 23 Jun 2022 13:21
 Operator : JC/MD
 Sample : N3202-02 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15

Quant Time: Jun 24 05:38:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270492	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	453099	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	396449	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	184202	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	144295	40.299	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	80.600%
35) Dibromofluoromethane	5.391	113	130854	44.316	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.640%
50) Toluene-d8	8.653	98	509969	47.276	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.560%
62) 4-Bromofluorobenzene	11.079	95	190607	46.856	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.720%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.959	59	52561	60.341	ug/l #	93
16) Acetone	2.386	43	3350	2.347	ug/l	93
19) Methyl tert-butyl Ether	3.117	73	57015	5.993	ug/l	95
22) Diisopropyl ether	3.764	45	3835	0.438	ug/l #	84
31) Cyclohexane	5.471	56	6512	1.355	ug/l #	89
39) Methylcyclohexane	7.391	83	2667	0.535	ug/l #	84
40) Benzene	6.044	78	272846	23.032	ug/l	99
52) Toluene	8.720	92	60125	7.926	ug/l	97
67) Ethyl Benzene	10.195	91	549750	39.285	ug/l	100
68) m/p-Xylenes	10.305	106	468966	85.934	ug/l	99
69) o-Xylene	10.647	106	86357	16.017	ug/l	96
73) Isopropylbenzene	10.964	105	49159	3.559	ug/l	97
78) n-propylbenzene	11.305	91	85437	5.492	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	136503	11.736	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	606625	52.293	ug/l	99
95) Naphthalene	13.774	128	264112	20.441	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062322\
Data File : VX029739.D
Acq On : 23 Jun 2022 13:21
Operator : JC/MD
Sample : N3202-02 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-15

Quant Time: Jun 24 05:38:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

