

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062322\
 Data File : VX029742.D
 Acq On : 23 Jun 2022 14:31
 Operator : JC/MD
 Sample : N3202-10DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22DL

Quant Time: Jun 24 05:39:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	268834	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	450439	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	379467	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	170224	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	144313	40.553	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.100%
35) Dibromofluoromethane	5.391	113	132262	45.057	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.120%
50) Toluene-d8	8.653	98	500416	46.664	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.320%
62) 4-Bromofluorobenzene	11.079	95	178348	44.101	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.200%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.977	59	53200	61.628	ug/l #	94
19) Methyl tert-butyl Ether	3.117	73	174006	18.403	ug/l	97
40) Benzene	6.050	78	13252	1.125	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	30916	2.884	ug/l	99
95) Naphthalene	13.780	128	15770	1.321	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062322\
Data File : VX029742.D
Acq On : 23 Jun 2022 14:31
Operator : JC/MD
Sample : N3202-10DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22DL

Quant Time: Jun 24 05:39:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

