

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062322\
 Data File : VX029744.D
 Acq On : 23 Jun 2022 15:17
 Operator : JC/MD
 Sample : N3202-11RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-25RE

Quant Time: Jun 24 05:39:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	268290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	446003	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	380858	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	169891	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	144069	40.567	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.140%
35) Dibromofluoromethane	5.391	113	128825	44.323	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.640%
50) Toluene-d8	8.653	98	494629	46.583	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.160%
62) 4-Bromofluorobenzene	11.085	95	178886	44.674	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.340%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.959	59	85443	105.036	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	322785	34.207	ug/l	97
22) Diisopropyl ether	3.763	45	20452	2.356	ug/l	91
80) 1,3,5-Trimethylbenzene	11.451	105	2800	0.261	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	7536	0.704	ug/l	97
95) Naphthalene	13.780	128	3117	0.262	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062322\
Data File : VX029744.D
Acq On : 23 Jun 2022 15:17
Operator : JC/MD
Sample : N3202-11RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-25RE

Quant Time: Jun 24 05:39:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

