

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062322\
 Data File : VX029746.D
 Acq On : 23 Jun 2022 16:04
 Operator : JC/MD
 Sample : N3317-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Jun 24 05:39:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 20 01:31:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	256564	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	424135	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	368600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	155046	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140049	41.237	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.480%
35) Dibromofluoromethane	5.391	113	125246	45.313	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.620%
50) Toluene-d8	8.653	98	478323	47.370	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.740%
62) 4-Bromofluorobenzene	11.079	95	171124	44.939	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.880%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX062322\
Data File : VX029746.D
Acq On : 23 Jun 2022 16:04
Operator : JC/MD
Sample : N3317-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Jun 24 05:39:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061822W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 20 01:31:01 2022
Response via : Initial Calibration

